Aptima® Multitest Swab Specimen Collection Kit

Patient collection procedure guide*

For COVID testing with nasal swab specimens



In addition to the Aptima Multitest Swab, a nasal swab can be collected and placed in VTM, UTM, STM, saline or liquid Amies for COVID testing.







Wash hands before starting. If you have any questions about this procedure, please ask your healthcare provider.

Partially peel open swab package and remove swab. Do not touch the soft tip or lay the swab down. If the soft tip is touched, laid down, or dropped, discard and get a new Aptima Multitest Swab Specimen Collection Kit. **Hold swab, placing thumb and forefinger in the middle of swab shaft over black score line.** Do not hold the swab shaft below the score line.

Carefully insert the swab into the first nostril until resistance is met at the level of the turbinates (less than one inch into the nostril). Rotate the swab a few times against the nasal wall and remove from nostril.



Using the same swab, carefully insert the swab into the second nostril until resistance is met at the level of the turbinates (less than one inch into the nostril). Rotate the swab a few times against the nasal wall and remove from nostril.



While holding swab in hand, unscrew tube cap. Do not spill tube contents. If tube contents are spilled, discard and replace with a new Aptima Multitest Swab Specimen Collection Kit.





Immediately place swab into transport tube so black score line is at top of tube. Align the score line with top edge of tube and carefully break shaft.





Discard top portion of shaft. Tightly screw cap onto tube. Return tube as instructed by your healthcare provider.

Specimens in the Aptima Multitest tube may be stored at 2°C to 30°C up to 6 days.

hologic.com | diagnostic.solutions@hologic.com | 1.888.484.4747

DS-10006-001 Rev. 001 © 2020 Hologic, Inc. All rights reserved. Hologic, Aptima and associated logos are trademarks and/or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries. This information is intended for medical professionals and is not intended as a product solicitation or promotion where such activities are prohibited. Because Hologic materials are distributed through websites, eBroadcasts and tradeshows, it is not always possible to control where such materials appear. For specific information on what products are available for sale in a particular country, please contact your Hologic representative or write to diagnostic.solutions@hologic.com.



^{*} Hologic provides this collection procedure guide as a general informational tool only; it is not an affirmative instruction or guarantee of performance. It is the sole responsibility of the clinician to read and understand the appropriate package insert and comply with applicable local, state and federal rules and regulations.